

FULL FINAL LABORATORY REPORT

LAB NUMBER: UH 01346520

SUBMITTED BY:
J.H ESTERHUIZEN
EH Makana Municipality
Medical Officer
Po Box 176
Grahamstown
Eastern Cape
6140

SAMPLE: ALICE DALE
SOURCE: DRINKING WATER
Sample Ref: PH19/1192
Client Ref No: FINAL RESERVOIR
Outbreak Case No:
Collected: 25/06/2019 10:30
Received: 26/06/2019 08:00
Printed: 27/06/2019 14:25

REPORT TO:
J.H ESTERHUIZEN
EH Makana Municipality
Medical Officer
Po Box 176
Grahamstown
Eastern Cape
6140

PUBLIC HEALTH

Specimen received: Water
Tests requested: Potable water

Potable Water:
Coliform count (Colilert) 0 per 100ml
E coli count (Colilert) 0 per 100ml

WATER DOMESTIC
ESSENTIAL MICROBIOLOGICAL CRITERIA AS PER SANS 241
Total bacterial count.....<=1000 per mL
Total coliforms.....<=10 per 100mL
Escherichia coli.....Absent in 100mL

The above test result relates only to the item/sample as received

REFERENCE METHOD(S):
Coliform Count - Colilert Method.....Colilert SANS 5521
E.coli Count - Colilert Method.....Colilert SANS 5521

Authorised by: D Bezuidenhout (Medical Technologist) Potable water

-- End of Laboratory Report --

FULL FINAL LABORATORY REPORT

LAB NUMBER: UH 01346517

SUBMITTED BY:
JH ESTERHUIZEN
EH Makana Municipality
Medical Officer
Po Box 176
Grahamstown
Eastern Cape
6140

SAMPLE: EAST FINAL RESEVOIR
SOURCE: DRINKING WATER

Sample Ref: PH19/1191
Client Ref No: RIEBEECK EAST
Outbreak Case No:
Collected: 25/06/2019 10:30
Received: 26/06/2019 08:00
Printed: 27/06/2019 14:25

REPORT TO:
JH ESTERHUIZEN
EH Makana Municipality
Medical Officer
Po Box 176
Grahamstown
Eastern Cape
6140

PUBLIC HEALTH

Specimen received: Water
Tests requested: Potable water

Potable Water:

Coliform count (Colilert) 0 per 100ml
E coli count (Colilert) 0 per 100ml

WATER DOMESTIC
ESSENTIAL MICROBIOLOGICAL CRITERIA AS PER SANS 241
Total bacterial count.....<=1000 per mL
Total coliforms.....<=10 per 100mL
Escherichia coli.....Absent in 100mL

The above test result relates only to the item/sample as received

REFERENCE METHOD(S):

Coliform Count - Colilert Method.....Colilert SANS 5521
E.coli Count - Colilert Method.....Colilert SANS 5521

Authorised by: D Bezuidenhout (Medical Technologist) Potable water

-- End of Laboratory Report --